

IPS TESTING

Test Report May 10, 2021 Page 1 of 6 SGS-IPS 00788-21

Report to: Tim Benjamin

MMX Manufacturing LLC 6508 NW 77th Court Miami, FL 33166

Sample Description: **Project IBR JN24302 (One Mask Sample)**

Date Received: April 23, 2021

Test(s) Requested: Flammability (16 CFR Part 1610), Resistance to Blood Penetration

(ASTM F 1862)

PO Number: SGS IBR JN24302

Analysis of One Mask Sample

SGS-IPS Testing performed the testing listed above on one mask sample provided by MMX Manufacturing LLC. The results are summarized in Tables 1 and 2 on the following pages of this report.

If you have any questions, please contact us.

SGS does not verify authenticity of any Brand/Trademark of products. Buyers must check if the product is genuine with the Brand/Trademark owner directly.

Authorized by

Jennifer Goggans Laboratory Manager Signed

Sean Walsh Lab Technician Physical Testing 920-749-3040

Project No. IBR JN24302

Table 1. Physical Properties

Medical Face Mask (JN24302)

Flammability (16 CFR Part 1610)





AT-1659

In Scope of ISO Y 17025

Accreditation (Y/N)

Orientation Nose to Chin

Test Side Outside

Burn Time (sec)

1 DNI
2 DNI
3 DNI
4 DNI
5 DNI
Average DNI

Conditioning Oven

Classifications Indicates Class 1.

IBE indicates the specimen Ignited But Extinguished.

DNI indicates the specimen Did Not Ignite.

Average calculated based on combusted samples only.

Table 2. Resistance to Blood Penetration (ASTM F 1862)

Medical Face Mask (JN24302)





AT-1659

In Scope of ISO Y 17025 Accreditation (Y/N)

Resistance to Blood Penetration (Pass/Fail)

1	Pass
2	Pass
3	Pass
4	Pass
5	Pass
6	Pass
7	Pass
8	Pass
9	Pass
10	Fail
11	Pass
12	Pass
13	Pass
14	Pass
15	Pass
16	Pass
17	Pass
18	Fail
19	Pass
20	Pass
21	Pass
22	Pass
23	Pass
24	Pass
25	Pass
26	Pass
27	Pass
28	Fail
29	Pass
30	Pass
31	Pass
32	Pass

Resistance to 29 Pass/3 Fail

Blood Penetration

Summary

Test Parameters

Synthetic Blood Lot 218201

#

Test Pressure 160

(mmHg)



Project No. IBR JN24302

Method(s) and Notes:

All valid results are included in the statistical analyses.

Revisions of SGS-IPS methods when used are current at the time of testing.

Samples tested and conditioned in TAPPI standard conditions unless requested otherwise by customer or otherwise specified.

Samples were not preconditioned.

16 CFR Ch. II Part 1610 (1-1-18 Edition) - Standard for the Flammability of Clothing Textiles Testing conducted as a Plain Surface Textile Fabric with modification IPS-T049.

If samples were oven-conditioned, samples were then cooled in a desiccator for a minimum of 15 minutes.

ASTM F 1862/F 1862M-17 Standard Test Method for Resistance of Medical Face Masks to Penetration by Synthetic Blood (Horizontal Projection of Fixed Volume at a Known Velocity)

A 2 mL check was performed at the beginning, middle and end of each sample.

Synthetic blood is purchased from Johnson, Moen & Co. Surface Tension is not independently verified after receipt and unused synthetic blood is stored in original plastic bottles.

Samples were conditioned prior to testing. Chamber conditions can be found as reported.

Analyzed by: <u>SW</u> Quality review by: <u>EB</u>

Date(s) of testing: May 7-10, 2021

Room Conditions

	Relative Humidity (%)	Temperature (°F)
Conditioning Environment	49.9	73.6
Maximum during testing Minimum during testing	49.9 49.9	73.8 73.6

Chamber Conditions ASTM F 1862 160mmHg

	Relative Humidity (%)	Temperature (°F)
Chamber Maximum Chamber Minimum	83.2 83.1	72.1 72.0

Note: See the method(s) cited above for available estimates of measurement uncertainty. Unless otherwise noted, sampling was performed by customer.



Report to MMX Manufacturing LLC SGS-IPS 00788-21

Project No. IBR JN24302

May 10, 2021 Page 5 of 6

This document is issued by the Company with the expectation that the document will be used internally only and not be used for any marketing or any other public use and is subject to its General Conditions of Service printed overleaf, available on request or accessible at www.sgs.com/en/Terms-and_conditions.htm and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at www.sgs.com/en/Terms-and-Conditions/terms-e-document.aspx. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Project No. IBR JN24302 Appendix A



Figure 1. Login Sample Photos Medical Face Mask (JN24302)

Product information is provided by applicant without verification of authentication of the brand.